Fuzion™
greater fit, function & freedom℠

NEW! AFO Options

The next generation of Fuzion designs provides lower profile options for patients with challenging bony architecture. Orthomerica's 2 new Fuzion AFO designs have a new low-profile plastic inner shell trim option.

#1. The addition of a semi-rigid composite strut to this design allows for dorsiflexion but resists plantarflexion.

#2. Rigid composite strut resists both dorsiflexion and plantarflexion. The unique needs of each patient will determine the optimal design choices.





KEY FEATURES

#1 Low Profile Plastic Trim (Polypropylene)

Leg size determines Polypropylene thickness. Low trim lines avoid bony prominences.

#2 Carbon Composite Posterior Strut

The posterior strut will be wider for the rigid option and narrower for the semi-rigid. The length of the strut will be from the bottom of the gastrocnemius muscle belly to the metatarsal heads.

ORTHOMERICA®
www.orthomerica.com/fuzion
877-737-8444

EC REP Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands

©2019 Orthomerica Products, Inc. All rights reserved. SL00254 Rev. B
6333 N Orange Blossom Trl, Orlando FL 32810

Date _____

ORTHOMERICA® FUZION® AFO 0585

6333 North Orange Blossom Trail, Orlando FL 32810 • www.orthomerica.com • phone 877-737-8444 • fax 877-737-8445

CALL TECHNICAL SUPPORT FOR ANY FABRICATION QUESTIONS

PATIENT INFO (PHI)

Last Name / ID _____ First Name _____

Gender
○ Male ○ Female

Age _____ Height _____ Weight _____

Diagnosis _____

Remarks

SHIPPING INFORMATION

Practitioner _____ Phone/Fax _____

Facility _____ PO Number _____

Shipping Company
○ UPS ○ FedEx ○ Other: _____

Service
○ Ground ○ 2 Day Air ○ Overnight

Need by _____

Ship to Address _____

City _____ State _____ Zip _____

Bill to Address _____

City _____ State _____ Zip _____


Shape Acquisition Via: ○ Cast ○ Scan | Affected Side: ○ Left ○ Right ☐ Bilateral

ALIGNMENT *MUST BE COMPLETED TO PREVENT DELAY OF ORDER*

Ankle Alignment (Dorsiflexion - Plantarflexion)
☐ Correct to _____ degrees ☐ Do not correct (Cast alignment OK)


Hindfoot Alignment
☐ Correct to _____ degrees ☐ Do not correct (Cast alignment OK)

Forefoot Alignment
Circle drawing below to indicate finished forefoot alignment



_____ _____ Include met height to floor in inches _____ _____

DORSAL EXTENSION -to control forefoot



☐ Full Length Tongue

Straps
☐ Add Toe Strap*
☐ Dacron Reinforced Straps*

Strap Color
○ White Default
○ Color or Transfer: _____


*Additional charges apply

Inner Liner Color
_____

Outer Foam Skin Color
○ Black ○ White


External Posting
○ No External Posting Default
○ Heel Posting*
○ Heel & Midfoot Posting*
○ Entire External Posting*
☐ Forefoot posting to balance any Supination or Pronation*
☐ Non Skid Bottom*

Measurements




If no measurements are provided, fabrication will be anatomical to mold.

Standard trim



Low profile trim



Standard Trim Plastic Options
○ Heat Adjustable Default
○ ProFlex®*
○ Copolymer
○ PolyPro®

Low Profile Trim Polypropylene Options
○ Semi-rigid Carbon Composite Posterior Strut*
○ Rigid Carbon Composite Posterior Strut*

©2019 Orthomerica Products, Inc. All Rights Reserved. FM002.90 Rev. C